



EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 027625 0080 Rev. 01

Manufacturer:

TRB Chemedica AG

Otto-Lilienthal-Ring 26
85622 Feldkirchen/München
GERMANY

SRN Manufacturer - DE-MF-000006919

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s). The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result.

The quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis.

The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. The surveillance assessment shall also include an assessment of the technical documentation for the device or devices concerned on the basis of further representative samples. All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with.

For details and certificate validity see: [www.tuvsud.com/ps-cert?q=cert:G10 027625 0080 Rev. 01](http://www.tuvsud.com/ps-cert?q=cert:G10_027625_0080_Rev_01)

Report No.:	713293304
Preceding Certificate No.:	G10 027625 0080 Rev. 00
Valid from:	2024-07-23
Valid until:	2029-01-16
Date of Initial Issuance:	2024-01-17

Christoph Dicks
Head of Certification/Notified Body

Issue date: 2024-07-23



EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
 (Class IIa and Class IIb Devices)

No. G10 027625 0080 Rev. 01

Classification: Class IIb
Device Group: Q02030301 - OPHTHALMOLOGY, HYALURONIC ACID AND DERIVATIVES
Intended Purpose: For sustained lubrication of the eyes
 Lubricant eye drops
 For the lubrication of contact lenses during wear

Classification: Class IIa
Device Group: Q02030301 - OPHTHALMOLOGY, HYALURONIC ACID AND DERIVATIVES
Intended Purpose: Viscoelastic solution for ocular surgery

The validity of this certificate depends on conditions and/or is limited to the following: -

Revision History:

Rev.	Dated	Report	Description
00	2024-01-17	713277096	- Initial issuance
01	2024-07-23	713293304	Supplemented: Device(s)/group of device(s) added